



ISO 9001/EN 29001/BS 5750
APPROVED BY BVQI

KYORITSU

NEW



(Actual Size)



Carrying Case

CARD TYPE DIGITAL MULTIMETER

- Display : 4000 counts
- Auto range.
- Diode test feature.
- Capacitance test feature.
- Continuity test.
- Automatically turns power off in about 15 minutes to conserve battery life.
- Designed to international safety standard IEC61010-1 CAT.Ⅲ 300V.

Specifications

DC V	400mV / 4 / 40 / 400 / 600V (Input Impedance 10MΩ) ±0.8%rdg±5dgt (400mV / 4 / 40 / 400V) ±1.0%rdg±5dgt (600V)
AC V	4 / 40 / 400 / 600V (Input Impedance 10MΩ) ±1.3%rdg±5dgt (4 / 40V) ±1.6%rdg±5dgt (400 / 600V)
Hz	10 / 100Hz / 1 / 10 / 100 / 1000kHz / 10MHz
Ω	400Ω / 4 / 40 / 400kΩ / 4 / 40MΩ ±1.0%rdg±5dgt (400Ω / 4 / 40 / 400kΩ / 4MΩ) ±2.5%rdg±5dgt (40MΩ)
Diode Test	4V Release Voltage : Approx. 0.4mA Test Current
Capacitance Test	4nF / 40nF / 400nF / 4μF / 40μF / 200μF
Continuity buzzer	400Ω (Buzzer sounds below 50Ω)
Withstand Voltage	3700V AC for 1 minute
Standard	IEC 61010-1 CAT.Ⅲ 300V IEC 61010-2-031 IEC 61326-1
Power Source	LR44(1.5V)×2
Dimensions	107(L)×54(W)×10(D)mm
Weight	70g approx.
Accessories	LR44×2 Carrying Case Instruction Manual

CARD TYPE DIGITAL MULTIMETER

MODEL 1017



Safety Warnings

Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely for correct use.
Failure to follow the safety rules can cause fire, trouble, electrical shock, etc.
Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

*The contents of this catalogue may be subject to change without notice.

■DISTRIBUTOR



KYORITSU ELECTRICAL INSTRUMENTS WORKS,LTD.

No. 5-20 Nakane 2-chome, Meguro-Ku, Tokyo, 152-0031 Japan
Phone :81-3-3723-0131 Fax:81-3-3723-0152
URL : <http://www.kew-ltd.co.jp>
E-mail : info@kew-ltd.co.jp
Factories:Uwajima & Ehime